

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Effective: 10/10/2022

Version: 3.0

Instrument #: 100422

Received From: JBFR

Date: 7/26/24

Analyst: C. OBryant

Reason for instrument return:

operator screen unresponsive despite
in the field troubleshooting.

^{COB 7/26/24}
Change instrument and ~~case~~ location in JusticeTrax Breath Alcohol Database

Ensure change in instrument status form is in JusticeTrax

Upload

Update agency information and polling status in DMHost Calling Program

Check for VOCs

From Intoximeters

Intoximeters Repair Paperwork (attached)