Alaska Scientific Crime Detection Laboratory Change in Instrument Status Form

Version: 6.0	ege in Instrument Status Form Effective: 1/17/2024
Date: 8/13/2025	Instrument #: 100675
Instrument Location: AST/Delta	
Supervisor Name and Agency: Carlie Gla	nister - SCDL
experienced 4 internal standard error 6/30/2025 passed. The instrument the 7/21/2025, and 7/28/2025. No breath agency. This instrument was effective	h the Instrument: 2025. Prior to it being put into service the instrument s on VOCs, 3 on 6/18/2025 and 1 on 6/19/2025. The VOC on en had internal standard errors on VOCs on 7/18/2025, a tests were conducted on the instrument while it was at the ly not in service, and is being deemed "Out of Service" on Service Change In Instrument Status Form (100675-2025-07-14-
If instrument was placed in service or re	emoved from service, fill out the appropriate sections below.
In Service Date/Time: NA	
Out of Service Date/Time: 06/30/2025	
For Use by SCDL	Additional Notes
Click or tap here to enter text.	C

Email completed form to dps.scdl.toxicology@alaska.gov.

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor



RE: Replacement Datamaster

From Stuart, Tyler M (DPS) <tyler.stuart@alaska.gov>

Date Mon 7/28/2025 9:08 AM

To Glaister, Carlie K (DPS) <carlie.glaister@alaska.gov>

I appreciate the prompt response. We'll keep an eye out for it. Also, I will also be more timely on the Change in Instrument Status Forms for the next swap. Thank you.

Respectfully,

-

SERGEANT TYLER STUART

ALASKA STATE TROOPERS
POST SUPERVISOR
"D" DETACHMENT: DELTA JUNCTION

DISPATCH: 907-895-4800

Fax: 907-895-4026

From: Glaister, Carlie K (DPS) <carlie.glaister@alaska.gov>

Sent: Monday, July 28, 2025 8:25 AM

To: Stuart, Tyler M (DPS) <tyler.stuart@alaska.gov>

Subject: Re: Replacement Datamaster

Dont worry about the in service Change In Instrument Status Form, we wound up filling one out on the agencies behalf using the date of the passing VOC.

I think we will need to replace the instrument. I will get the ball rolling on that and we will send tracking info shortly.

Carlie Glaister

Forensic Scientist 3 State of Alaska Scientific Crime Detection Laboratory 907-269-5689



From: Stuart, Tyler M (DPS) < tyler.stuart@alaska.gov>

Sent: Monday, July 28, 2025 7:49 AM

To: Glaister, Carlie K (DPS) < <u>carlie.glaister@alaska.gov</u>>

Cc: DPS SCDL Toxicology (DPS sponsored) < dps.scdl.toxicology@alaska.gov

Subject: RE: Replacement Datamaster

Good morning,

I completely missed this email. I will get you the new form here shortly. Additionally, the replacement Datamaster failed another VOC. I tried to turn it off for a bit and try the VOC again. It didn't work this time. It presented a diagnostic check failure. Do you have any other guidance? Thanks.

Respectfully,

SERGEANT TYLER STUART

ALASKA STATE TROOPERS POST SUPERVISOR

"D" DETACHMENT: DELTA JUNCTION

DISPATCH: 907-895-4800

Fax: 907-895-4026

From: Glaister, Carlie K (DPS) <carlie.glaister@alaska.gov>

Sent: Tuesday, July 1, 2025 1:07 PM

To: Stuart, Tyler M (DPS) < tyler.stuart@alaska.gov >

Subject: RE: Replacement Datamaster

Can I also get a form for 100675, the new one, going in to service? Thank you so much!

Carlie Glaister
Forensic Scientist 3
State of Alaska Scientific Crime Detection Laboratory
907-269-5689



From: Stuart, Tyler M (DPS) < tyler.stuart@alaska.gov >

Sent: Tuesday, July 1, 2025 12:59 PM

To: Glaister, Carlie K (DPS) < <u>carlie.glaister@alaska.gov</u>> **Cc:** Walton, Derek J (DPS) < <u>derek.walton@alaska.gov</u>>

Subject: RE: Replacement Datamaster

No problem! The form is attached; the original will be sent with the instrument. First time filling one of those out, let me know if I need to add or change anything thanks.

Respectfully,

SERGEANT TYLER STUART

ALASKA STATE TROOPERS
POST SUPERVISOR

"D" DETACHMENT: DELTA JUNCTION

DISPATCH: 907-895-4800

Fax: 907-895-4026

From: Glaister, Carlie K (DPS) <carlie.glaister@alaska.gov>


Sent: Tuesday, July 1, 2025 9:51 AM

To: Stuart, Tyler M (DPS) < tyler.stuart@alaska.gov> **Cc:** Walton, Derek J (DPS) < derek.walton@alaska.gov>

Subject: RE: Replacement Datamaster

Also Tyler, can you please send me an in service change in instrument status form at your earliest convenience? Thank you!

Carlie Glaister
Forensic Scientist 3
State of Alaska Scientific Crime Detection Laboratory 907-269-5689



From: Stuart, Tyler M (DPS) < tyler.stuart@alaska.gov >

Sent: Tuesday, July 1, 2025 8:43 AM

To: Glaister, Carlie K (DPS) < <u>carlie.glaister@alaska.gov</u>> **Cc:** Walton, Derek J (DPS) < <u>derek.walton@alaska.gov</u>>

Subject: RE: Replacement Datamaster

Good morning,

Thank you both for the assistance. The replacement Datamaster appears to be functioning properly now; I received a passing VOC upon my return from leave.

I withheld sending our previous instrument until this was resolved. My admin returns Monday; she will send it out then. Let me know if you need any other action or information on my part. Thank you again.

Respectfully,

SERGEANT TYLER STUART

ALASKA STATE TROOPERS
POST SUPERVISOR
"D" DETACHMENT: DELTA JUNCTION
DISPATCH: 907-895-4800

Fax: 907-895-4026

From: Glaister, Carlie K (DPS) < carlie.glaister@alaska.gov>

Sent: Friday, June 27, 2025 10:09 AM

To: Stuart, Tyler M (DPS) <tyler.stuart@alaska.gov>

Subject: FW: Replacement Datamaster

Hello Tyler - I will be working in the breath lab next week. Derek passed the info on to me. It looks like the internal standard is reading persistently high and hasn't had a passing VOC. Unfortunately I don't have much advice for a quick and easy in the field fix for this. Since it looked normal right before leaving the lab I'm assuming that the filter wheel might have being impacted by shipping. When you are back from leave can I have you give it one more attempt since its been a while to see if its figured itself out, and if you get the same error we will ship you a new instrument.

Carlie Glaister
Forensic Scientist 3
State of Alaska Scientific Crime Detection Laboratory
907-269-5689



From: Walton, Derek J (DPS) <derek.walton@alaska.gov>

Sent: Friday, June 27, 2025 10:03 AM

To: Glaister, Carlie K (DPS) < carlie.glaister@alaska.gov >

Subject: FW: Replacement Datamaster

AST Delta Junction email chain about DataMaster issues with 100675

Derek J Walton

Forensic Scientist Alaska State Crime Lab 907-269-5628

From: DPS SCDL Toxicology (DPS sponsored) < dps.scdl.toxicology@alaska.gov

Sent: Monday, June 23, 2025 10:14 AM

To: Walton, Derek J (DPS) < derek.walton@alaska.gov>

Subject: FW: Replacement Datamaster

Derek J Walton

Forensic Scientist Alaska State Crime Lab 907-269-5628

From: Stuart, Tyler M (DPS) < tyler.stuart@alaska.gov >

Sent: Thursday, June 19, 2025 6:53 AM

To: DPS SCDL Toxicology (DPS sponsored) < dps.scdl.toxicology@alaska.gov>

Subject: RE: Replacement Datamaster

Good morning,

Reference the replacement Datamaster, I can not seem to get it to pass the Diagnostic Check. I've tried 3 VOC's so far. I don't have any Datamaster Supervisors available at the moment to consult. Do you have any guidance to get ours up and running? Thanks.

Respectfully,

SERGEANT TYLER STUART

ALASKA STATE TROOPERS POST SUPERVISOR

"D" DETACHMENT: DELTA JUNCTION

DISPATCH: 907-895-4800

Fax: 907-895-4026

From: DPS SCDL Toxicology (DPS sponsored) <dps.scdl.toxicology@alaska.gov>

Sent: Wednesday, June 18, 2025 2:45 PM

Subject: RE: Replacement Datamaster

Thank you!

Derek J Walton

Forensic Scientist Alaska State Crime Lab 907-269-5628

From: Stuart, Tyler M (DPS) < tyler.stuart@alaska.gov>

Sent: Wednesday, June 18, 2025 12:12 PM

To: DPS SCDL Toxicology (DPS sponsored) < dps.scdl.toxicology@alaska.gov>

Cc: Lentz, Angela A (DPS) < angela.lentz@alaska.gov>

Subject: Replacement Datamaster

Good afternoon,

Attached is the change in instrument status for the Datamaster previously at Delta AST Post. The DMT will be sent out early next week. Angela Lentz will reply all with shipping information. Thank you.

Respectfully,

SERGEANT TYLER STUART

ALASKA STATE TROOPERS
POST SUPERVISOR
"D" DETACHMENT: DELTA JUNCTION

DISPATCH: 907-895-4800

Fax: 907-895-4026