

# Alaska Scientific Crime Detection Laboratory

## DataMaster Calibration Form

Issued: 6/11/2014  
Effective: 6/11/2014

Version: DMCAF 2014 R0  
Status: Active

Instrument # 100350

Date: 9/14/17

### Reason for Calibration

Interference Detected + Internal Std error in field.

#### 1. Wet Bath Simulator Solution

- a. Level 0.100
- b. Manufacturer Guth
- c. Lot 16160
- d. Expiration Date 7/19/18

#### 2. Wet Bath Simulator

- a. Serial Number DR6267
- b. Date Calibrated 3/27/17

#### 3. Attach Calibration Factors

CALIBRATION FACTORS

-----  
Alaska Department of Public Safety  
DATAMASTER dmt: 100350  
-----

Date: 09/14/2017  
Time: 12:31:53

OPERATOR NAME:  
BRANDI BARNETT  
CERTIFICATE #: 7038

Ca = 0.1000  
CAL = 0.982587    0.800 <= CAL < 1.200  
b1 = 0.0008    0.0000 <= b1 < 0.0040  
b2 = 0.0040    0.0010 <= b2 < 0.0100  
b3 = 0.0003    0.0000 <= b3 < 0.0040  
Xq = 0.0987    0.0500 <= Xq < 0.2000  
a21 = 1.151125    1.050 <= a21 < 1.300  
a31 = 0.482170    0.300 <= a31 < 0.800