

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 4.0

Effective: 8/10/2022

Date: 10/28/2023

Instrument #: 100667

Instrument Location: Anchorage Jail

Supervisor Name and Agency: Jason Cusack APD # 9120

Describe the Status Change or Issue with the Instrument:

I was notified the Datamaster was having the same issue as yesterday where it was not recognizing when a breath sample was being provided. I removed the instrument from service to return it to the crime lab.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: [Click or tap here to enter text.](#)

Out of Service Date/Time: 10/28/2023 0023

For Use by SCDL

Additional Notes

[Click or tap here to enter text.](#)

Email completed form to dps.scdl.toxicology@alaska.gov
For questions contact the Breath Alcohol Program at 907-269-5740

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor

From: Cusack, Jason <jason.cusack@anchorageak.gov>
Sent: Saturday, October 28, 2023 12:32 AM
To: DPS SCDL Toxicology (DPS sponsored) <dps.scdl.toxicology@alaska.gov>
Subject: RE: status chang 100667

Well that didn't actually fix the issue like I thought as Worland had the same issue I did. I have removed 100667 from service and moved 100687 to take it's place in the North processing room. I will bring the instrument in next week.

Jason



Jason Cusack

Officer – Impaired Driving Enforcement Unit

Anchorage Police Department
716 W 4th Ave, Anchorage, AK 99501
Office: (907) 786-8703

www.anchoragepolice.com

