

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 4.0

Effective: 8/10/2022

Date: 4/7/2023

Instrument #: 100423

Instrument Location: 3025 Montgomery Rd, Fort Wainwright, AK 99703

Supervisor Name and Agency: LT Julayha, Sazahan Fort Wainwright Police Department

Describe the Status Change or Issue with the Instrument:

Out of Service due to numerous/continuous errors.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: [Click or tap here to enter text.](#)

Out of Service Date/Time: 04/07/2023 at 1400

For Use by SCDL

Additional Notes

On 04/03/2023, we were notified by Breath Test Supervisor Julayha Sazahan that a subject spit into the breath hose through the mouthpiece during a breath test. Since then, the instrument has continuously received Ambient Error and Standard out of Range Status messages with an external standard reading of 0.071. The instrument was pulled out of service to be sent back to the lab for calibration or repair. TLM 04/10/2023

Email completed form to dps.scdl.toxicology@alaska.gov
For questions contact the Breath Alcohol Program at 907-269-5740

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor