

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100695

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 09/02/2016

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.65: 0.079  
LOT #: AG608504  
EXPIRATION: 03/25/2018  
TANK PRESSURE: 1199 psi

BLANK TEST	0.000	21:50
INTERNAL STANDARD	VERIFIED	21:50
EXTERNAL STANDARD	0.080	21:50
BLANK TEST	0.000	21:51
EXTERNAL STANDARD	0.079	21:51
BLANK TEST	0.000	21:52
EXTERNAL STANDARD	0.080	21:53
BLANK TEST	0.000	21:54
EXTERNAL STANDARD	0.080	21:54
BLANK TEST	0.000	21:55
EXTERNAL STANDARD	0.080	21:55
BLANK TEST	0.000	21:56

Average = 0.0798  
Std Dev = 0.0004

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 49.1°C PASSED  
Breath Tube = 46.7°C PASSED

### PUMP INFO

Flow Rate = 4.226 L/M PASSED

### DETECTOR INFO

PUMP ON PASSED  
PUMP OFF PASSED

### FILTER INFO

Filter 1 PASSED  
Filter 2 PASSED  
Filter 3 PASSED

INTERNAL STANDARD PASSED

I, Charles R. Foster, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

*Charles R. Foster* 9/16/16

Charles R. Foster  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 16<sup>th</sup> day of Sept, 2016

*Nikki Roth*  
Nikki Roth, Notary Public  
My Commission Expires With Office



CRB 9/16/16