

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100689

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 08/17/2021

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.67: 0.079  
LOT #: AG113402  
EXPIRATION: 05/14/2023  
TANK PRESSURE: 1110 psi

BLANK TEST	0.000	15:49
INTERNAL STANDARD	VERIFIED	15:50
EXTERNAL STANDARD	0.079	15:50
BLANK TEST	0.000	15:51
EXTERNAL STANDARD	0.079	15:51
BLANK TEST	0.000	15:52
EXTERNAL STANDARD	0.079	15:52
BLANK TEST	0.000	15:53
EXTERNAL STANDARD	0.079	15:53
BLANK TEST	0.000	15:54
EXTERNAL STANDARD	0.079	15:54
BLANK TEST	0.000	15:55

Average = 0.0790  
Std Dev = 0.0000

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.07  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 49.0°C	PASSED
Breath Tube = 40.6°C	PASSED

### PUMP INFO

Flow Rate = 3.861 L/M	PASSED
-----------------------	--------

### DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

### FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------

File only: Instrument was never placed in service at Galena due to leaking regulator.