

	State of Alaska Scientific Crime Detection Laboratory		
	DataMaster Calibration Form – DMCAF2011R1		
	Issuing Authority – Nita Bolz, Scientific Director		
Approval date: 03/10/11	Effective date: 03/10/11	Archive date: Active	

**DataMaster DMT Calibration Form**

Instrument # 102360

Date: 4/12/11

1. Wet Bath Simulator Solution

- a. Level 0.100
- b. Manufacturer Buth
- c. Lot 10260
- d. Expiration Date 10/25/12

2. Wet Bath Simulator

- a. Serial Number DR 4161
- b. Date Calibrated 8/23/10

3. Attach Calibration Factors

Breath Alcohol Analyst C. O' Bryant

*COB/b*

CALIBRATION FACTORS

-----  
Alaska Department of Public Safety  
DATAMASTER dmt: 100360  
-----

Date: 04/12/2011  
Time: 13:43:41

SUPERVISOR NAME:  
COLLEEN S OBRYANT  
CERTIFICATE #: 6237

LOT #: GUTH100  
BOTTLE #: 10260  
EXPIRATION: ~~11/25/2012~~ 10/25/12  
COB 4/13/11

Ca = 0.1000	
CAL = 0.977991	0.800 <= CAL < 1.200
b1 = 0.0002	0.0000 <= b1 < 0.0040
b2 = 0.0066	0.0010 <= b2 < 0.0100
b3 = 0.0002	0.0000 <= b3 < 0.0040
Xq = 0.1225	0.0500 <= Xq < 0.2000
a21 = 1.220006	1.050 <= a21 < 1.300
a31 = 0.503914	0.300 <= a31 < 0.800

COB 4/12