

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 03/18/2017 Instrument Serial Number: 100696

Instrument Location: Yakutat Police Dept.; Yakutat AK 99689

Supervisor Name and Agency: None / Yakutat PD.

Describe the Status Change or Issue with the Instrument:

Instrument needed to be replaced. Replaced with
instrument # 100373.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 03/18/2017 11:45 Hrs.

For Use by SCDL

Additional Notes

Instrument went out of service due to an expired
tank on 2/5/17. When new tank was being installed
instrument gave gas flow errors and was replaced.

SMB 7/11/17

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740