

Alaska Scientific Crime Detection Laboratory  
Change in Instrument Status Form

Version: 6.0

Effective: 1/17/2024

Date: 4/24/2024

Instrument #: 100392

Instrument Location: Homer PD

Supervisor Name and Agency: Carlie Glaister - SCDL

Describe the Status Change or Issue with the Instrument:

During In Field Instrument Review it was noticed that a subject was tested 8 times in a row which included 2 pump errors, 2 blank errors, and 3 incomplete status messages followed by a completed test. The breath test supervisor, Taylor Crowder was contacted and notified that it is not advisable to test a subject more than 3 times and the recommendation is to get a blood sample if a completed test is not provided within 3 attempts. See attached email.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: NA

Out of Service Date/Time: NA

-----  
**For Use by SCDL**

Additional Notes

Click or tap here to enter text.

**Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov).**

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor

## Glaister, Carlie K (DPS)

---

**From:** Glaister, Carlie K (DPS)  
**Sent:** Wednesday, April 24, 2024 8:48 AM  
**To:** Crowder, Taylor (DPS sponsored)  
**Subject:** DataMaster 100392

Hello Taylor!

While reviewing status messages on DataMaster 100392 I came across a subject who was tested 8 times on 3/1/2024. A completed test occurred after 7 tests that included 2 pump errors, 2 blank errors and 3 incomplete status messages. We do not advise testing a subject more than 3 times. It would be recommended to get a blood draw if unable to get a completed test within 3 attempts. Please pass this information along to your operators.

100392	3/1/2024 9:03 AM	Homer PD	A...		Pump Error
100392	3/1/2024 9:26 AM	Homer PD	A...		
100392	3/1/2024 9:32 AM	Homer PD	A...		Blank Error
100392	3/1/2024 9:36 AM	Homer PD	A...		
100392	3/1/2024 9:44 AM	Homer PD	A...		Pump Error
100392	3/1/2024 9:45 AM	Homer PD	A...		
100392	3/1/2024 9:55 AM	Homer PD	A...		Blank Error
100392	3/1/2024 9:58 AM	Homer PD	A...	0.27100000	

Please let me know if you have any questions,

*Carlie Glaister*

Forensic Scientist 3  
State of Alaska Scientific Crime Detection Laboratory  
907-269-5689