

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100351

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 05/24/2024

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.54: 0.079  
LOT #: AG310901  
EXPIRATION: 04/19/2025  
TANK PRESSURE: 1213 psi

BLANK TEST	0.000	19:28
INTERNAL STANDARD	VERIFIED	19:28
EXTERNAL STANDARD	0.078	19:28
BLANK TEST	0.000	19:29
EXTERNAL STANDARD	0.078	19:29
BLANK TEST	0.000	19:30
EXTERNAL STANDARD	0.078	19:30
BLANK TEST	0.000	19:31
EXTERNAL STANDARD	INTERFERENCE	19:31

Blank Error

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 49.1°C	PASSED
Breath Tube = 44.8°C	PASSED

### PUMP INFO

Flow Rate = 4.499 L/M	PASSED
-----------------------	--------

### DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

### FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

### INTERNAL STANDARD

PASSED

File only DM 10/03/24