

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100375

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 09/12/2020

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 28.00: 0.075
LOT #: AG018102
EXPIRATION: 06/29/2022
TANK PRESSURE: 1184 psi

BLANK TEST	0.000	09:09
INTERNAL STANDARD	VERIFIED	09:09
EXTERNAL STANDARD	0.074	09:09
BLANK TEST	0.000	09:10
EXTERNAL STANDARD	0.074	09:10
BLANK TEST	0.000	09:11
EXTERNAL STANDARD	0.072	09:11
BLANK TEST	0.000	09:12
EXTERNAL STANDARD	0.071	09:12
BLANK TEST	0.000	09:13
EXTERNAL STANDARD	0.072	09:13

Blank Error

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C	PASSED
Breath Tube = 45.4°C	PASSED

PUMP INFO

Flow Rate = 4.464 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------

File only JMR 09/16/20