

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Issued: 6/11/2014  
Effective: 6/11/2014

Version: CIISF 2014 R0  
Status: Active

Date: 1/27/15 Instrument Serial Number: 100687

Instrument Location: BPD

Supervisor Name and Agency: Joe Corbett / BPD

Describe the Status Change or Issue with the Instrument:

REPEATED STATUS MESSAGES / TROUBLE SHOOTING NOT SUCCESSFUL

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: \_\_\_\_\_

Out of Service Date/Time: 1/27/15 1315 Hrs

For Use by SCDL	Additional Notes

Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)

For questions contact the Breath Alcohol Program at 907-269-5740