



State of Alaska
 Department of Public Safety
Scientific Crime Detection Laboratory

4805 Dr. Martin Luther King Jr. Ave.
 Anchorage, Alaska 99507
 David Kanaris, Chief, Forensic Laboratories



Breath Alcohol Calibration Report
Calibration Item: DataMaster DMT
Serial Number: 100354

Report Issue Date: 05/22/2024
Date Received: 2024-03-08
Condition of Item: No apparent damage

Method Used: DataMaster DMT Calibration / Current Breath Alcohol Procedure Manual
Calibration Range: 0.020 g/210L - 0.300 g/210L

As-Found Data								
Nominal Values (g/210L)	Target Values (g/210L)	Lot Number	Results (g/210L)					Mean (g/210L)
0.020	0.020	AG217902	0.020	0.020	0.020	0.020	0.020	0.020
0.080	0.079	AG310901	0.077	0.077	0.078	0.077	0.077	0.077
0.300	0.298	AG327804	0.294	0.294	0.294	0.294	0.294	0.294

As Found - Evaluation Analyst: Carlie Glaister
 As Found Conditions - Barometric Pressure (inHg): 29.60 Barometer Used: Fisher S/N 230307229 Temperature (C): 20.1 Relative Humidity (%): 25
 As-Found Lab Activity Date Range: 03/28/2024 - 03/29/2024

Replaced O-Ring.

As-Left Data									
Nominal Values (g/210L)	Target Values (g/210L)	Lot Number	Results (g/210L)					Mean (g/210L)	Uncertainty (g/210L)
0.020	0.020	AG217902	0.020	0.020	0.019	0.020	0.020	0.020	+/- 0.002
0.040	0.040	AG217904	0.039	0.039	0.039	0.039	0.039	0.039	+/- 0.003
0.080	0.080	AG335202	0.078	0.078	0.078	0.078	0.078	0.078	+/- 0.004
0.100	0.100	AG333401	0.098	0.098	0.099	0.099	0.099	0.099	+/- 0.004
0.150	0.149	AG332004	0.146	0.145	0.145	0.145	0.145	0.145	+/- 0.005
0.300	0.299	AG327804	0.295	0.295	0.295	0.296	0.296	0.295	+/- 0.011

As Left Conditions - Barometric Pressure (inHg): 29.77 Barometer Used: Fisher S/N 230307229 Temperature (C): 21.9 Relative Humidity (%): 21
 As-Left Lab Activity Date Range: 05/16/2024 - 05/16/2024

Calibration is performed using dry gas standards traceable to NIST. Standards are sourced from an external provider accredited to ISO/IEC 17025 and ISO 17034 standards. Internal instrument barometer readings are traceable to NIST. External barometers are calibrated by an external calibration provider accredited to the ISO/IEC 17025 standard.

The As-Left results listed above include the estimated measurement uncertainty with a coverage factor k=2 and an expanded coverage probability of 95%. The estimated uncertainty applies to the calibration process and not to subject breath samples.

This report applies only to the item being calibrated and shall not be reproduced, except in full, without written approval from the laboratory.

Results authorized by Derek Walton on 05/16/2024

derek.walton@alaska.gov

Technically and administratively reviewed by Colleen O'Bryant on 05/22/2024

Alaska Scientific Crime Detection Laboratory

DataMaster Calibration Form

Effective: 1/23/2024

Version: 4.0

Instrument #: 100354 Analyst or Technician Name: Derek Watten
DataMaster DMT Calibration Start Date: 05/16/24

Diagnostic Test

- Perform Diagnostic Test (View Details)
- Internal Standard Value \leq 3.00%

Technician Screen and Calibration Factors

- Adjust Voltages (if needed)
 - Lamp (2.60 V Maximum)
 - Bias
 - Cooler (2.00 V Maximum)
 - Chopper Frequency
 - Detector Voltage (-0.100V to 0.100V)
- Set RFI Sensitivity
- Adjust internal barometer reading to match NIST-traceable barometer
Barometer/Thermometer/Hygrometer S/N: 230307229
Calibration Due Date: 05/02/25
- Verify Volume Setting (1.50 L +/- 0.10 L), adjust if needed
- Print Technician Screen
- Print "Cal Factors"

Minimum Volume Test

- Perform Minimum Volume (Subject) Test: 1.5 liters (1.3 L to 1.7 L)

As Left Linearity Testing

Barometric Pressure (in Hg) 29.77 Temperature(C) 21.9 Relative Humidity(%) 21
Approved Concentrations (g/210L): 0.020, 0.040, 0.080, 0.100, 0.150, 0.300

Functional Tests

- Incomplete
- Invalid
- Interference (Volatiles) - *Print with View Details*
- Maximum Attempts Exceeded
- Radio Frequency Interference (RFI)

1. Acetone Reagent Lot Number: 20240501KS Expiration Date: 08/01/24
Simulator S/N: SD1088 Thermometer Verification Performed?
2. Isopropanol Reagent Lot Number: 20240501MNC Expiration Date: 08/01/24
Simulator S/N: DR416 Thermometer Verification Performed?
3. Negative (Ethanol Only) Lot Number: 224741201 Expiration Date: 08/24/24 (0.100 solution)
Simulator S/N: SD177 Thermometer Verification Performed?

Calibrated External Thermometer S/N: 230631106
Calibration Due Date: 09/19/25

Administrative

- Verify Stylus Calibration
- Reset Options
- Clear Memory
- Verification of Calibration (VOC)
- Perform Non-Drinking Subject Test
- Upload Instrument
- Remove Dry Gas Tank Before Shutdown
- Add Ready for Tech Review Note

Completion

Lab Work End Date: 05/16/24
Data Interpretation Date (Technician): NA (if applicable)
Data Interpretation Date (Analyst): 05/16/24

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor

DIAGNOSTIC RECORD

Alaska Department of Public Safety
DATAMASTER dmt: 100354

Date: 05/16/2024
Time: 09:23:16

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C PASSED
(44.0°C - 52.0°C)
Breath Tube = 48.1°C PASSED
(38.0°C - 50.0°C)

SETTINGS

Lamp Voltage = 1.64 V PASSED
(0.95 V - 2.23 V)
Cooler Voltage = 1.53 V PASSED
(0.95 V - 2.21 V)
Bias Voltage = 78 V PASSED
(47 V - 111 V)
Chopper Freq = 548 Hz PASSED
(475 Hz - 575 Hz)
Barometer = 29.8 in

PUMP INFO

Flow Rate = 4.755 L/M PASSED
(3.500 L/M - 6.500 L/M)

DETECTOR INFO

PUMP ON PASSED
(0.003 v <= 0.010 v)
PUMP OFF PASSED
(0.002 v <= 0.010 v)

FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD ✓ PASSED
Xq = 0.097 1.65% (0.00% - 4.00%)

Technician Screen

DATAMASTER dmt: 100354

Date: 05/16/2024
Time: 9:43 AM

Temperatures (°C)

Sample Cell: 48.9°C
Breath Tube: 47.4°C

Current Barometer 29.8 in
Volume (Ltr) 1.57 ✓

Settings

Lamp: 1.75 V ✓
Bias: 80 V ✓
Cooler: 1.56 V ✓
Chopper: 525 Hz

Voltages (V)

Flow: 1.05 V
Detector: -0.082 V ✓

CALIBRATION FACTORS

Alaska Department of Public Safety
DATAMASTER dmt: 100354

Date: 09/27/2011
Time: 13:04:23

OPERATOR NAME:
COLLEEN S OBRYANT
CERTIFICATE #: 6237

LOT #: GUTH100
BOTTLE #: 11050
EXPIRATION: 04/06/2013

Ca = 0.1000
CAL = 0.969216 0.800 <= CAL < 1.200
b1 = 0.0001 0.0000 <= b1 < 0.0040
b2 = 0.0049 0.0010 <= b2 < 0.0100
b3 = 0.0001 0.0000 <= b3 < 0.0040
Xq = 0.0953 0.0500 <= Xq < 0.2000
a21 = 1.175978 1.050 <= a21 < 1.300
a31 = 0.450287 0.300 <= a31 < 0.800

SUBJECT TEST

Alaska Department of Public Safety
DATAMASTER dmt: 100354

Date: 05/16/2024
Time: 09:45:39

SUBJECT NAME:
TEST TEST
LICENSE #: XXXX

OPERATOR NAME:
DEREK WALTON
OPERATOR'S #: 9508


DEPT/AGENCY:
MINIMUM VOLUME
CASE/REPORT: XXXX
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.79: 0.080
LOT #: AG335202
EXPIRATION: 12/18/2025

DIAGNOSTIC CHECK	PASSED	09:48
BLANK TEST	0.000	09:48
INTERNAL STANDARD	VERIFIED	09:48
EXTERNAL STANDARD	0.078	09:49
BLANK TEST	0.000	09:49
SUBJECT SAMPLE (Vol=1.55L)	0.000	09:50
BLANK TEST	0.000	09:51
EXTERNAL STANDARD	0.078	09:52
BLANK TEST	0.000	09:52

DW



LINEARITY CHECK


Alaska Department of Public Safety
 DATAMASTER dmt: 100354

Date: 05/16/2024
 Time: 10:14:26

SUPERVISOR NAME:
 DEREK WALTON
 OPERATOR'S #: 9508

TARGET	AVERAGE	STD. DEV.
0.020	0.0198	0.0004
0.040	0.0390	0.0000
0.080	0.0780	0.0000
0.100	0.0986	0.0005
0.149	0.1452	0.0004
0.299	0.2954	0.0005

SLOPE = 0.988054
 R² = 0.999938



DM

LINEARITY CHECK: Control Test 1

Alaska Department of Public Safety
 DATAMASTER dmt: 100354






Date: 05/16/2024
 Time: 10:14:29

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.020
 TARGET AT 29.78: 0.020
 LOT #: AG217902
 EXPIRATION: 06/28/2024
 TANK PRESSURE: 473 psi

BLANK TEST	0.000	10:15
INTERNAL STANDARD	VERIFIED	10:16
EXTERNAL STANDARD	0.020	10:16
BLANK TEST	0.000	10:17
EXTERNAL STANDARD	0.020	10:17
BLANK TEST	0.000	10:18
EXTERNAL STANDARD	0.019	10:18
BLANK TEST	0.000	10:19
EXTERNAL STANDARD	0.020	10:19
BLANK TEST	0.000	10:20
EXTERNAL STANDARD	0.020	10:20
BLANK TEST	0.000	10:21

Average = 0.0198
 Std Dev = 0.0004



LINEARITY CHECK: Control Test 2

Alaska Department of Public Safety
 DATAMASTER dmt: 100354






Date: 05/16/2024
 Time: 10:32:10

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.040
 TARGET AT 29.78: 0.040
 LOT #: AG217904
 EXPIRATION: 06/28/2024
 TANK PRESSURE: 711 psi

BLANK TEST	0.000	10:33
INTERNAL STANDARD	VERIFIED	10:33
EXTERNAL STANDARD	0.039	10:33
BLANK TEST	0.000	10:34
EXTERNAL STANDARD	0.039	10:34
BLANK TEST	0.000	10:35
EXTERNAL STANDARD	0.039	10:36
BLANK TEST	0.000	10:36
EXTERNAL STANDARD	0.039	10:37
BLANK TEST	0.000	10:38
EXTERNAL STANDARD	0.039	10:38
BLANK TEST	0.000	10:39

Average = 0.0390
 Std Dev = 0.0000



LINEARITY CHECK: Control Test 3

Alaska Department of Public Safety
DATAMASTER dmt: 100354






Date: 05/16/2024
Time: 10:50:22

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.78: 0.080
LOT #: AG335202
EXPIRATION: 12/18/2025
TANK PRESSURE: 1183 psi

BLANK TEST	0.000	10:51
INTERNAL STANDARD	VERIFIED	10:51
EXTERNAL STANDARD	0.078	10:52
BLANK TEST	0.000	10:52
EXTERNAL STANDARD	0.078	10:53
BLANK TEST	0.000	10:54
EXTERNAL STANDARD	0.078	10:54
BLANK TEST	0.000	10:55
EXTERNAL STANDARD	0.078	10:55
BLANK TEST	0.000	10:56
EXTERNAL STANDARD	0.078	10:56
BLANK TEST	0.000	10:57

Average = 0.0780
Std Dev = 0.0000



LINEARITY CHECK: Control Test 4

Alaska Department of Public Safety
 DATAMASTER dmt: 100354






Date: 05/16/2024
 Time: 11:00:19

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.100
 TARGET AT 29.77: 0.100
 LOT #: AG333401
 EXPIRATION: 11/30/2025
 TANK PRESSURE: 1023 psi

BLANK TEST	0.000	11:01
INTERNAL STANDARD	VERIFIED	11:02
EXTERNAL STANDARD	0.098	11:02
BLANK TEST	0.000	11:03
EXTERNAL STANDARD	0.098	11:03
BLANK TEST	0.000	11:04
EXTERNAL STANDARD	0.099	11:04
BLANK TEST	0.000	11:05
EXTERNAL STANDARD	0.099	11:05
BLANK TEST	0.000	11:06
EXTERNAL STANDARD	0.099	11:06
BLANK TEST	0.000	11:07

Average = 0.0986
 Std Dev = 0.0005



LINEARITY CHECK: Control Test 5

Alaska Department of Public Safety
DATAMASTER dmt: 100354






Date: 05/16/2024
Time: 11:10:50

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.150
TARGET AT 29.77: 0.149
LOT #: AG332004
EXPIRATION: 11/16/2025
TANK PRESSURE: 594 psi

BLANK TEST	0.000	11:11
INTERNAL STANDARD	VERIFIED	11:12
EXTERNAL STANDARD	0.146	11:12
BLANK TEST	0.000	11:13
EXTERNAL STANDARD	0.145	11:13
BLANK TEST	0.000	11:14
EXTERNAL STANDARD	0.145	11:14
BLANK TEST	0.000	11:15
EXTERNAL STANDARD	0.145	11:15
BLANK TEST	0.000	11:16
EXTERNAL STANDARD	0.145	11:16
BLANK TEST	0.000	11:17

Average = 0.1452
Std Dev = 0.0004



LINEARITY CHECK: Control Test 6






Alaska Department of Public Safety
 DATAMASTER dmt: 100354

Date: 05/16/2024
 Time: 11:20:29

EXTERNAL STANDARD INFORMATION
 NOMINAL: 0.300
 TARGET AT 29.77: 0.299
 LOT #: AG327804
 EXPIRATION: 10/05/2025
 TANK PRESSURE: 145 psi

BLANK TEST	0.000	11:21
INTERNAL STANDARD	VERIFIED	11:21
EXTERNAL STANDARD	0.295	11:22
BLANK TEST	0.000	11:22
EXTERNAL STANDARD	0.295	11:23
BLANK TEST	0.000	11:24
EXTERNAL STANDARD	0.295	11:24
BLANK TEST	0.000	11:25
EXTERNAL STANDARD	0.296	11:25
BLANK TEST	0.000	11:26
EXTERNAL STANDARD	0.296	11:26
BLANK TEST	0.000	11:27

Average = 0.2954
 Std Dev = 0.0005



SUBJECT TEST

Alaska Department of Public Safety
DATAMASTER dmt: 100354

Date: 05/16/2024
Time: 12:16:56


SUBJECT NAME:
TEST TEST
LICENSE #: XXXX

OPERATOR NAME:
DEREK WALTON
OPERATOR'S #: 9508

DEPT/AGENCY:
INCOMPLETE
CASE/REPORT: XXXX
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 29.77: 0.080
LOT #: AG335202
EXPIRATION: 12/18/2025

DIAGNOSTIC CHECK PASSED 12:17
BLANK TEST 0.000 12:18
INTERNAL STANDARD VERIFIED 12:18
EXTERNAL STANDARD 0.078 12:18
BLANK TEST 0.000 12:19
SUBJECT SAMPLE INCOMPLETE 12:21



SUBJECT TEST

Alaska Department of Public Safety
DATAMASTER dmt: 100354

Date: 05/16/2024
Time: 12:30:54

SUBJECT NAME:
TEST TEST
LICENSE #: XXXX

OPERATOR NAME:
DEREK WALTON
OPERATOR'S #: 9508


DEPT/AGENCY:
MAXIMUM ATTEMPTS
CASE/REPORT: XXXX
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 29.78: 0.080
LOT #: AG335202
EXPIRATION: 12/18/2025

DIAGNOSTIC CHECK PASSED 12:32
BLANK TEST 0.000 12:32
INTERNAL STANDARD VERIFIED 12:32
EXTERNAL STANDARD 0.078 12:32
BLANK TEST 0.000 12:33
SUBJECT SAMPLE XXX 12:34

Maximum Attempts Exceeded

DM



SUBJECT TEST

Alaska Department of Public Safety

DATAMASTER dmt: 100354

Date: 05/16/2024

Time: 12:37:19

SUBJECT NAME:

TEST TEST

LICENSE #: XXXX

OPERATOR NAME:

DEREK WALTON

OPERATOR'S #: 9508

DEPT/AGENCY:

INVALID

CASE/REPORT: XXXX

TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080

TARGET AT 29.78: 0.080

LOT #: AG335202

EXPIRATION: 12/18/2025

DIAGNOSTIC CHECK

PASSED 12:38

0.000 12:38

INTERNAL STANDARD

VERIFIED 12:38

0.078 12:38


EXTERNAL STANDARD

0.000 12:39

INVALID 12:40

SUBJECT SAMPLE

Handwritten signature



SUBJECT TEST

Alaska Department of Public Safety
DATAMASTER dmt: 100354

Date: 05/16/2024
Time: 13:20:03

SUBJECT NAME:
TEST TEST
LICENSE #: XXXX


OPERATOR NAME:
DEREK WALTON
OPERATOR'S #: 9508

DEPT/AGENCY:
RADIO FREQUENCY
CASE/REPORT: XXXX
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 29.76: 0.080
LOT #: AG335202
EXPIRATION: 12/18/2025

DIAGNOSTIC CHECK PASSED 13:24
BLANK TEST 0.000 13:24
INTERNAL STANDARD VERIFIED 13:24
EXTERNAL STANDARD 0.078 13:24
BLANK TEST 0.000 13:25

RFI DETECTED



SUBJECT TEST

Alaska Department of Public Safety
DATAMASTER dmt: 100354

Date: 05/16/2024
Time: 13:33:44

SUBJECT NAME:
TEST TEST
LICENSE #: XXXX

OPERATOR NAME:
DEREK WALTON
OPERATOR'S #: 9508

DEPT/AGENCY:
INTERFERENCE ACETONE
CASE/REPORT: XXXX
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 29.77: 0.080
LOT #: AG335202
EXPIRATION: 12/18/2025

DIAGNOSTIC CHECK	PASSED	13:34
BLANK TEST	0.000	13:34
INTERNAL STANDARD	VERIFIED	13:34
Xq = 0.097 (1.47%)		
EXTERNAL STANDARD	0.078	13:35
X[1] = 0.0775 (-0.0016) (0.0024)		
BLANK TEST	0.000	13:36
SUBJECT SAMPLE	XXX	13:36
X[1] = 0.1100 (-0.0203) (0.0009)		

Interference Detected

Filter Lavage: -0.0012

Filter 3 average: 0.0025 dm

05/16/24



SUBJECT TEST

Alaska Department of Public Safety
DATAMASTER dmt: 100354

Date: 05/16/2024
Time: 13:39:23

SUBJECT NAME:
TEST TEST
LICENSE #: XXXX


OPERATOR NAME:
DEREK WALTON
OPERATOR'S #: 9508

DEPT/AGENCY:
INTERFERENCE ISOPROPANOL
CASE/REPORT: XXXX
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 29.77: 0.080
LOT #: AG335202
EXPIRATION: 12/18/2025

DIAGNOSTIC CHECK	PASSED	13:40
BLANK TEST	0.000	13:40
INTERNAL STANDARD	VERIFIED	13:40
Xq = 0.097 (1.93%)		
EXTERNAL STANDARD	0.078	13:40
X[1] = 0.0784 (-0.0008) (0.0027)		
BLANK TEST	0.000	13:41
SUBJECT SAMPLE	XXX	13:42
X[1] = 0.1187 (-0.0262) (0.0022)		

Interference Detected



SUBJECT TEST

Alaska Department of Public Safety
DATAMASTER dmt: 100354

Date: 05/16/2024
Time: 13:45:10

SUBJECT NAME:
TEST TEST
LICENSE #: XXXX


OPERATOR NAME:
DEREK WALTON
OPERATOR'S #: 9508

DEPT/AGENCY:
INTERFERENCE NEGATIVE
CASE/REPORT: XXXX
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 29.77: 0.080
LOT #: AG335202
EXPIRATION: 12/18/2025

DIAGNOSTIC CHECK	PASSED	13:46
BLANK TEST	0.000	13:46
INTERNAL STANDARD	VERIFIED	13:47
Xq = 0.097 (1.65%)		
EXTERNAL STANDARD	0.078	13:47
X[1] = 0.0778 (-0.0010)	(0.0026)	
BLANK TEST	0.000	13:48
SUBJECT SAMPLE (Vol=2.06L)	0.095	13:48
X[1] = 0.0954 (-0.0002)	(0.0013)	
BLANK TEST	0.000	13:49
EXTERNAL STANDARD	0.077	13:49
X[1] = 0.0773 (-0.0012)	(0.0024)	
BLANK TEST	0.000	13:50

DM



SUBJECT TEST

Alaska Department of Public Safety
DATAMASTER dmt: 100354

Date: 05/16/2024
Time: 14:44:07

SUBJECT NAME:
TEST TEST
LICENSE #: XXXX

OPERATOR NAME:
DEREK WALTON
OPERATOR'S #: 9508

DEPT/AGENCY:
NON-DRINKING SUBJECT
CASE/REPORT: XXXX
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 29.76: 0.080
LOT #: AG335202
EXPIRATION: 12/18/2025

DIAGNOSTIC CHECK	PASSED	14:45
BLANK TEST	0.000	14:45
INTERNAL STANDARD	VERIFIED	14:45
EXTERNAL STANDARD	0.078	14:45
BLANK TEST	0.000	14:46
SUBJECT SAMPLE (Vol=2.69L)	0.000	14:46
BLANK TEST	0.000	14:47
EXTERNAL STANDARD	0.079	14:47
BLANK TEST	0.000	14:48

WAL

