

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100344

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 02/22/2018

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 30.63: 0.082  
LOT #: AG735001  
EXPIRATION: 12/16/2019  
TANK PRESSURE: 1189 psi

BLANK TEST	0.000	11:46
INTERNAL STANDARD	VERIFIED	11:46
EXTERNAL STANDARD	0.081	11:46
BLANK TEST	0.000	11:47
EXTERNAL STANDARD	0.081	11:47
BLANK TEST	0.000	11:48
EXTERNAL STANDARD	0.082	11:48
BLANK TEST	0.000	11:49
EXTERNAL STANDARD	0.081	11:49
BLANK TEST	0.000	11:50
EXTERNAL STANDARD	0.081	11:51
BLANK TEST	0.000	11:51

Average = 0.0812  
Std Dev = 0.0004

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.02  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 49.1°C PASSED  
Breath Tube = 47.7°C PASSED

### PUMP INFO

Flow Rate = 6.093 L/M PASSED

### DETECTOR INFO

PUMP ON PASSED  
PUMP OFF PASSED


### FILTER INFO

Filter 1 PASSED  
Filter 2 PASSED  
Filter 3 PASSED

INTERNAL STANDARD PASSED


I, Charles R. Foster, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

  
Charles R. Foster  
Scientific Director  
State Breath Alcohol Program

4/25/18

Subscribed and sworn before me this 25 day of 04, 20 18

  
Carlie K. Bailey, Notary Public  
My Commission Expires With Office



Supervised By  
COB 4/24/18  
JMW 04/24/18