

# Alaska Scientific Crime Detection Laboratory

## DataMaster Check In Form

Effective: 3/20/2025

Version: 4.0

Instrument #: 100345

Received From: ST. Mary's AST

Date: 12/16/25

Analyst: Kaija Swisher

Reason for instrument return:

Filter wheel errors. Issues starting up. Unable to load ticket base.

### From the field

Change instrument and case location in JusticeTrax Breath Alcohol Database ☒

Ensure change in instrument status form is in JusticeTrax ☒

Upload ☒ (Unable to load ticket DB, modem disconnected) - uploaded via SD card.  
Update agency information and polling status in DMHost Calling Program ☒ Resolved ticket issue - KS  
Check for VOCs ☒ 12/16/25

### From Intoximeters

Intoximeters Repair Paperwork (attached) ☐

Change instrument and case location in JusticeTrax Breath Alcohol Database ☐

Reset Options and Clear Memory ☐