

CALIBRATION FACTORS

Alaska Department of Public Safety
DATAMASTER dmt: 100351

Date: 03/18/2020
Time: 08:09:54

OPERATOR NAME:
INTOX I METERS
CERTIFICATE #: 1

Ca	=	0.2000		
CAL	=	0.975058	0.800	<= CAL < 1.200
b1	=	0.0006	0.0000	<= b1 < 0.0040
b2	=	0.0055	0.0010	<= b2 < 0.0100
b3	=	0.0008	0.0000	<= b3 < 0.0040
Xq	=	0.1217	0.0500	<= Xq < 0.2000
a21	=	1.157872	1.050	<= a21 < 1.300
a31	=	0.464353	0.300	<= a31 < 0.800