

	State of Alaska Scientific Crime Detection Laboratory		
	DataMaster Calibration Form - DMCAF2011R1		
	Issuing Authority - Nita Bolz, Scientific Director		
	Approval date: 03/10/11	Effective date: 03/10/11	Archive date: Active

DataMaster DMT Calibration Form

Instrument # 100388

Date: 5/2/11

1. Wet Bath Simulator Solution

- a. Level 0.100
- b. Manufacturer Both
- c. Lot 10260
- d. Expiration Date 10/25/10

2. Wet Bath Simulator

- a. Serial Number DR-41601
- b. Date Calibrated 8/23/10

3. Attach Calibration Factors

Breath Alcohol Analyst C. O' Bryant

CALIBRATION FACTORS

Alaska Department of Public Safety
DATAMASTER dmt: 100388

Date: 05/02/2011
Time: 14:23:26

SUPERVISOR NAME:
COLLEEN S O'BRYANT
CERTIFICATE #: 6237

LOT #: GUTH100
BOTTLE #: 10260
EXPIRATION: 10/25/2012

Ca	=	0.1000		
CAL	=	0.991067	0.800	<= CAL < 1.200
b1	=	0.0003	0.0000	<= b1 < 0.0040
b2	=	0.0054	0.0010	<= b2 < 0.0100
b3	=	0.0002	0.0000	<= b3 < 0.0040
Xq	=	0.0956	0.0500	<= Xq < 0.2000
a21	=	1.153567	1.050	<= a21 < 1.300
a31	=	0.558289	0.300	<= a31 < 0.800

COB

Final page for report

\

\\ANCSCLSAN01\JTRAX\CASEDOCS\B-100388\B-100388_ECF-20140

2/19/2014