

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 8/20/15 Instrument Serial Number: 100360 BMB
100343 9/2/15

Instrument Location: PETERSBURG

Supervisor Name and Agency: CAPTAIN JOHN HAMILTON, PETERSBURG PD

Describe the Status Change or Issue with the Instrument:

INSTRUMENT SHOWED FATAL ERROR MESSAGE. RE-BOOED PERFORMED VOC AND NON-DRINKING SUBJECT TEST. INSTRUMENT NOW APPEARS TO BE WORKING NORMALLY

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: _____

For Use by SCDL

Additional Notes

Instrument 100343 was Petersburg PD's old instrument and it was not in service at this time. I checked DM Host and a VOC and non drinking subject test were run on 100360 (Petersburg PD's new instrument at in service at this time) on 8/21/15. The instrument number on the top of the form was changed to reflect the correct instrument #.

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740