

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Issued: 12/15/2017  
Effective: 12/15/2017

Version: CIISF 2017 R0  
Status: Active

Date: 7-18-21 Instrument Serial Number: 100411

Instrument Location: Fairbanks Police Department

Supervisor Name and Agency: Lane K. Bonham Fairbanks PD

### Describe the Status Change or Issue with the Instrument:

When another Officer attempted to run a subject test, he observed his subject actively blowing but was not getting a steady tone and it sounded like air was escaping inside the instrument. I witnessed the same thing when we had the subject try another test. I took the top cover off and conducted a test subject test on myself. When providing a sample of my breath, I could feel air coming out of the black air regulator. It appears the air flow regulator is stuck open and will not accept a subject test. Instrument will still complete External Standard tests with no issue.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: \_\_\_\_\_

Out of Service Date/Time: 7-16-21 at 0130 hours

**For Use by SCDL**

**Additional Notes**

---

---

---

---

---

---

---

---

---

---

Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)

For questions contact the Breath Alcohol Program at 907-269-5740

SUBJECT TEST

Alaska Department of Public Safety  
DATAMASTER dmt: 100411

Date: 07/16/2021  
Time: 01:02:28

SUBJECT NAME:  
TEST TEST  
LICENSE #: 00000000

OPERATOR NAME:  
LANE K BONHAM  
OPERATOR'S #: 9282

DEPT/AGENCY:  
FBK1  
CASE/REPORT: 00000  
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.52: 0.079  
LOT #: AG923401  
EXPIRATION: 08/22/2021

DIAGNOSTIC CHECK PASSED 01:03  
BLANK TEST 0.000 01:03  
INTERNAL STANDARD VERIFIED 01:03  
EXTERNAL STANDARD 0.078 01:03  
BLANK TEST 0.000 01:04  
SUBJECT SAMPLE INCOMPLETE 01:06

