

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100680

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 4/28/2015

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.26: 0.078  
LOT #: 11313080A1  
EXPIRATION: 6/1/2015  
TANK PRESSURE: 720 psi

BLANK TEST	0.000	08:44
INTERNAL STANDARD	VERIFIED	08:44
EXTERNAL STANDARD	0.077	08:44
BLANK TEST	0.000	08:45
EXTERNAL STANDARD	0.079	08:45
BLANK TEST	0.000	08:46
EXTERNAL STANDARD	0.078	08:47
BLANK TEST	0.000	08:47
EXTERNAL STANDARD	0.078	08:48
BLANK TEST	0.000	08:49
EXTERNAL STANDARD	0.078	08:49
BLANK TEST	0.000	08:50

Average = 0.0780  
Std Dev = 0.0007

## Diagnostic Check

### VERSIONS

DMT: 3.00  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.9°C PASSED  
Breath Tube = 47.7°C PASSED

### PUMP INFO

Flow Rate = 4.192 L/M PASSED

### DETECTOR INFO

PUMP ON PASSED  
PUMP OFF PASSED

### FILTER INFO


Filter 1 PASSED  
Filter 2 PASSED  
Filter 3 PASSED

INTERNAL STANDARD

PASSED

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.



  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

5/15/2015

Subscribed and sworn before me this 15<sup>th</sup> day of May, 2015



Nikki Roth, Notary Public  
My Commission Expires With Office



DMB  
4/28/15