

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 3/7/18 Instrument Serial Number: 100694

Instrument Location: Petersburg Police Department, 14 S Nordic Dr. Petersburg, AK 99833

Supervisor Name and Agency: Derek N. Thorsen, Petersburg Police Department

Describe the Status Change or Issue with the Instrument:

I initiated a tank change procedure on the device a received a (Diagnostic Check Failed) status message during VOC. I rebooted the device and it will no longer power-up. Power supplies are present and working properly.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 3/7/18, 1900 hours

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740