

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 2/2/2020 Instrument Serial Number: 100699

Instrument Location: Anchorage Jail

Supervisor Name and Agency: Ofc J Roberts 7905, APD

Describe the Status Change or Issue with the Instrument:

The instrument has given several Gass Flow Error status messages. The instrument has been removed from service as of 2344 hrs until advised from Crime Lab.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 2/2/2020 2344 hrs

For Use by SCDL	Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 2/4/2020 Instrument Serial Number: 100699

Instrument Location: Anchorage Jail

Supervisor Name and Agency: Jason Cusack 9120, Anchorage Police

Describe the Status Change or Issue with the Instrument:

I attempted the run a VOC with the same bottle and received the same gas flow error. I changed out the bottles and completed a VOC which seemed to fix the problem. I then ran a test sample and the Datamaster is working normally again. It was put back into service at 0454 on 2/4/2020.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: 02/04/2020 0454

Out of Service Date/Time: _____

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 2-11-2020 Instrument Serial Number: B-100699

Instrument Location: Anchorage Jail / Room #B-104

Supervisor Name and Agency: P.E. Martin / APD

Describe the Status Change or Issue with the Instrument:

Gas Flow Error - Won't pass VOC (per officer Sitz)

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 2-7-2020 by officer L. Bowe

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740