

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 02/15/2019 Instrument Serial Number: 100685

Instrument Location: Kodiak PD

Supervisor Name and Agency: Derek Walton - SCDL

Describe the Status Change or Issue with the Instrument:

Communication error 1 and communication error 5 in field.
Pushed full version of 3.02 software remotely via Interactive mode.

Asked officer Thomas Harris (Kodiak PD) to reboot instrument.
Ran VOCs in field which resulted in Table interference.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: NA

Out of Service Date/Time: NA

For Use by SCDL

Additional Notes

Instrument will be replaced. JW 02/15/19

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740