

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100255

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 05/05/2021

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.40: 0.079  
LOT #: AG923401  
EXPIRATION: 08/22/2021  
TANK PRESSURE: 712 psi

BLANK TEST	0.000	16:38
INTERNAL STANDARD	VERIFIED	16:38
EXTERNAL STANDARD	0.076	16:39

Blank Error

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 2.05  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.8°C	PASSED
Breath Tube = 39.4°C	PASSED

### PUMP INFO

Flow Rate = 4.549 L/M	PASSED
-----------------------	--------

### DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

### FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

### INTERNAL STANDARD

PASSED

File only DMV 06/11/21