	State of Alaska Scientific Crime Detection Laboratory		
	DataMaster Calibration Form - DMCAF2011R1		
	Issuing Authority - Nita Bolz, Scientific Director		
Approval date: 03/10/11	Effective date: 03/10/11	Archive date: Active	

DataMaster DMT Calibration Form

Instrument # 100362

Date: 3/15/11

1. Wet Bath Simulator Solution

- a. Level 0.100
- b. Manufacturer Guth
- c. Lot 9380
- d. Expiration Date 12/17/2011

2. Wet Bath Simulator

- a. Serial Number DR4161
- b. Date Calibrated 8/23/10

3. Attach Calibration Factors

Breath Alcohol Analyst C. Bryant

CALIBRATION FACTORS


-----  
Alaska Department of Public Safety  
DATAMASTER dmt: 100362  
-----

Date: 03/15/2011  
Time: 07:01:28

SUPERVISOR NAME:  
COLLEEN S OBRYANT  
CERTIFICATE #: 6237

LOT #: GUTH100  
BOTTLE #: 9380  
EXPIRATION: 12/17/2011

Ca = 0.1000  
CAL = 0.977426    0.800 <= CAL < 1.200  
b1 = 0.0004    0.0000 <= b1 < 0.0040  
b2 = 0.0058    0.0010 <= b2 < 0.0100  
b3 = 0.0000    0.0000 <= b3 < 0.0040  
Xq = 0.0727    0.0500 <= Xq < 0.2000  
a21 = 1.161401    1.050 <= a21 < 1.300  
a31 = 0.532998    0.300 <= a31 < 0.800

	State of Alaska Scientific Crime Detection Laboratory		
	DataMaster Certification Form – DMCEF2011R0		
	Issuing Authority – Nita Bolz, Scientific Director		
	Approval date: 03/01/11	Effective date: 03/01/11	Archive date: Active

**DataMaster DMT Certification Form**Instrument # 100362Analyst Colleen O'BryantDate 3/7/11

1. Diagnostic Test (attached)
2. Voltage Checks
  - a. Adjust voltages, if needed 
    - i. Lamp (1.01 V dc – 2.37 V dc)
    - ii. Bias (48 V dc – 112 V dc)
    - iii. Cooler (1.18 V dc – 2.74 V dc)
    - iv. Chopper Frequency (475 hz – 575 hz)
  - b. Set RFI
  - c. Adjust barometer reading, if needed
  - d. Adjust volume setting, if needed
  - e. Print Technician Screen (attached)
3. Print Calibration Factors (attached)
4. Volume Test - Minimum \_\_\_\_\_ mL
5. Dry Gas Linearity Test (attached)
  - a. Levels Used 

i.	0.020g/210L Mnf _____	Lot _____	Exp _____
ii.	0.040g/210L Mnf _____	Lot _____	Exp _____
iii.	0.080g/210L Mnf _____	Lot _____	Exp _____
iv.	0.100g/210L Mnf _____	Lot _____	Exp _____
v.	0.200g/210L Mnf _____	Lot _____	Exp _____
vi.	0.300g/210L Mnf _____	Lot _____	Exp _____
6. Functional Tests (attached)
  - a. Incomplete
  - b. Invalid
  - c. Interference 
    - i. Acetone - Mnf \_\_\_\_\_ Lot \_\_\_\_\_ Level \_\_\_\_\_ Exp \_\_\_\_\_
      1. Simulator Calibration Date \_\_\_\_\_ SN \_\_\_\_\_
    - ii. Isopropanol - Mnf \_\_\_\_\_ Lot \_\_\_\_\_ Level \_\_\_\_\_ Exp \_\_\_\_\_
      1. Simulator Calibration Date \_\_\_\_\_ SN \_\_\_\_\_
  - d. RFI
7. Calibrate Stylus
8. Verify card reader is functional
9. Upload
10. Reset Options
11. Non-Drinking Subject Test (attached)

CALIBRATION FACTORS

-----  
Alaska Department of Public Safety  
DATAMASTER dmt: 100362  
-----

Date: 01/19/2011  
Time: 11:09:32

Ca = 0.0990  
CAL = 1.002040    0.800 <= CAL < 1.200  
b1 = 0.0005    0.0000 <= b1 < 0.0040  
b2 = 0.0056    0.0010 <= b2 < 0.0100  
b3 = 0.0002    0.0000 <= b3 < 0.0040  
Xq = 0.0724    0.0500 <= Xq < 0.2000  
a21 = 1.171025    1.050 <= a21 < 1.300  
a31 = 0.542539    0.300 <= a31 < 0.800

Technician Screen

-----  
DATAMASTER dmt: 100362  
-----

Date: 03/07/2011  
Time: 2:09 PM

Temperatures (°C)  
Sample Cell: 48.9°C  
Breath Tube: 46.6°C

Current Barometer 29.7 in  
Volume (Ltr) 1.57

Settings  
Lamp: 1.83 V  
Bias: 80 V  
Cooler: 1.61 V  
Chopper: 526 Hz

Voltages (V)  
Flow: 1.03 V  
Detector: 0.504 V