

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100685

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 12/08/2016

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 28.81: 0.077  
LOT #: 01315080A3  
EXPIRATION: 02/05/2017  
TANK PRESSURE: 602 psi

BLANK TEST	0.000	16:24
INTERNAL STANDARD	VERIFIED	16:24
EXTERNAL STANDARD	0.072	16:25
BLANK TEST	0.000	16:26
EXTERNAL STANDARD	0.072	16:26
BLANK TEST	0.000	16:27
EXTERNAL STANDARD	0.072	16:27
BLANK TEST	0.000	16:28
EXTERNAL STANDARD	0.072	16:28
BLANK TEST	0.000	16:29
EXTERNAL STANDARD	0.072	16:29
BLANK TEST	0.000	16:30

Average = 0.0720  
Std Dev = 0.0000

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.8°C PASSED  
Breath Tube = 48.1°C PASSED

### PUMP INFO

Flow Rate = 5.292 L/M PASSED

### DETECTOR INFO

PUMP ON PASSED  
PUMP OFF PASSED


### FILTER INFO

Filter 1 PASSED  
Filter 2 PASSED  
Filter 3 PASSED

INTERNAL STANDARD PASSED


I, Charles R. Foster, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


 12/23/16

Charles R. Foster  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 23<sup>rd</sup> day of Dec, 2016



Nikki Roth, Notary Public  
My Commission Expires With Office



BMB  
12/20/16