

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument
State of Alaska

Serial #: 100689

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 12/27/2019

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 28.90: 0.077
LOT #: AG826101
EXPIRATION: 09/18/2020
TANK PRESSURE: 1104 psi

BLANK TEST	0.000	22:51
INTERNAL STANDARD	VERIFIED	22:51
EXTERNAL STANDARD	0.076	22:52
BLANK TEST	0.000	22:52
EXTERNAL STANDARD	0.075	22:53
BLANK TEST	0.000	22:54
EXTERNAL STANDARD	0.077	22:54
BLANK TEST	0.000	22:55
EXTERNAL STANDARD	0.076	22:55
BLANK TEST	0.000	22:56
EXTERNAL STANDARD	0.076	22:56
BLANK TEST	0.000	22:57

Average = 0.0760
Std Dev = 0.0007

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C PASSED
Breath Tube = 44.4°C PASSED

PUMP INFO

Flow Rate = 3.889 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED


FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD PASSED


I, Brandi M. Barnett, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

 1/10/20

Brandi M. Barnett
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 10 day of Jan 20 20


Notary Public
My Commission Expires With Office



JM 12/30/19