

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 6.0

Effective: 1/17/2024

Date: 1/14/2025

Instrument #: 100352

Instrument Location: NSBPD Squad Room

Supervisor Name and Agency: Christopher Dominey

Describe the Status Change or Issue with the Instrument:

Out of Service. Screen will not respond to any pressure or touch

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: Click or tap here to enter text.

Out of Service Date/Time: 01/14/2025

For Use by SCDL

Additional Notes

The instrument was turned off/on and it did not resolve the issue. MNC 1/14/25

Email completed form to dps.scdl.toxicology@alaska.gov.

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor