

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100357

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date:03/15/2020

External Standard Test Values

Diagnostic Check

VERSIONS

DMT: 3.02

PIC: 3.02

Modem: 2.6

Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C

PASSED

Breath Tube = 42.1°C

PASSED

PUMP INFO

Flow Rate = 4.665 L/M

PASSED

DETECTOR INFO

PUMP ON

FAILED

PUMP OFF

PASSED

FILTER INFO

Filter 1

PASSED

Filter 2

PASSED

Filter 3

PASSED

INTERNAL STANDARD

PASSED

Diagnostic Check Failed