

# Alaska Scientific Crime Detection Laboratory

## DataMaster Check In Form

Issued: 7/10/2015  
Effective: 7/10/2015

Version: DMCIF 2015 R1  
Status: Active

Instrument S/N 100421

Received From NPAS

Date 8/3/15

Analyst Colleen O'Bryant

1. Reason for instrument return:

Returned from NPAS

2. Change instrument and case location in JusticeTrax Breath Alcohol Database

3. Ensure change in instrument status form is in JusticeTrax

4. Upload

5. Update agency information and polling status in DM Host Calling Program

6. Check for VOCs

From NPAS

7. NPAS Repair Paperwork (attached)

Work Ord No 49666

National Patent Analytical Systems, Inc.  
Service Action Report

Serial No 100421

Date 6/18/15

Customer Alaska Dept of Public

Phone 907-269-5582

Contact Bolz

War Exp 6/16/13

Mx Exp

Ship Damage?  Y  N

Cust Pick Up??  Y  N

Initial Comments: note any shipping damage or estimates required, along with accessories

Describe problem seen and/or attach documents received

internal std error

Keyword

Finding

Error Code

Enter Hours 2.0

Repair Comments and work Performed

Ran multiple filter tests with no errors. VOC test gave errors, updated software to correct. set voltages, ran all standard tests ok

Solution Lot No 15040

Voltage Set Values In				Voltage Set Values Out			
Lamp	1.630	Brth Vol	1.570	Lamp	1.750	Brth Vol	1.570
Meter	1.076	Filter Dif	0.422	Meter	0.283	Filter Dif	0.497
Stability	0.002			Stability	0.002		

Part No	Qty Ord	Pick	Back Ordered	Total shin	Description	Price Ea
9DMT	1.0				DATAMASTER DMT SERVICE RETURN 100421	
NOTE	1.0				NANUK CASE	
r500-1	1.0				Service Charge for Repair	0.000
r500-2	1.0				Service Charge for Repair, Additional Hours	0.000

Date Completed 7/8/15

Technician r. martin

Pick up Signature

**Packing List**

From: National Patent Analytical Systems Inc.  
2090 Harrington Mem. Rd., Mansfield, Ohio 44903

Cust No. AK6056A    Ord Date 6/18/15  
Sales Order No. 49666

Ship Date 7/14/15  
Packing List No. 74287

Cust Purch Ord No. COLLEEN

**Ship To:**

Alaska Dept of Public Safety  
Scientific Crime Detection  
4805 Dr. Martin Luther King Dr.

Anchorage                      Ak            99507

**Bill To:**

Alaska Dept of Public Safety  
Scientific Crime Detection  
4805 Dr Martin Luther King Dr.

Anchorage                      Ak            99507    **Freight**

25.00

**Attn:**

Ship Via FedEx

**Comments**    Service for 100421

Part / Drawing No / Rev No	Quantity Ordered	This Shipment	Back Ordered	Total Ship To Date	Description	NP Invent Ref
No Draw # / NONE 100421 Serial Nos. _	1.0	1		1.0	DATAMASTER DMT SERVICE RETURN	9DMT
No Draw # / NONE Serial Nos. _	1.0	1		1.0	NANUK CASE	NOTE
No Draw # / NONE Serial Nos. _	1.0	1		1.0	Service Charge for Repair	r500-1
No Draw # / NONE Serial Nos. _	1.0	1		1.0	Service Charge for Repair, Additional Hours	r500-2