

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100697

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 12/27/2022

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 28.76: 0.077
LOT #: AG113402
EXPIRATION: 05/14/2023
TANK PRESSURE: 174 psi

BLANK TEST	0.000	03:39
INTERNAL STANDARD	VERIFIED	03:39
EXTERNAL STANDARD	0.072	03:40
BLANK TEST	0.000	03:41
EXTERNAL STANDARD	0.071	03:41

Standard Out of Range

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.8°C	PASSED
Breath Tube = 45.8°C	PASSED

PUMP INFO

Flow Rate = 4.124 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------

File Only DMV 02/10/23