

Alaska Scientific Crime Detection Laboratory  
Change in Instrument Status Form

Version: 6.0

Effective: 1/17/2024

Date: 6/23/2025

Instrument #: 100363

Instrument Location: AST Nenana

Supervisor Name and Agency: Derek Walton – SCDL on the phone with Trooper Hollenbeck – AST Nenana

Describe the Status Change or Issue with the Instrument:

I viewed the instrument in Interactive Mode to determine why a VOC did not complete. The tank pressure was too low to initiate a VOC. I contacted Trooper Hollenbeck over the phone and requested a tank change be conducted. The tank change and VOC completed normally on 06/18/2025.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: NA – Tank change completed 06/18/2025

Out of Service Date/Time: NA

-----  
**For Use by SCDL**

Additional Notes

Click or tap here to enter text.

**Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov).**

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor