

# Alaska Scientific Crime Detection Laboratory

## DataMaster Evaluation Form

Issued: 7/10/2015  
Effective: 7/10/2015

Version: DMEF2015 R0  
Status: Active

Instrument # 100689

Date: 10/24/16

Analyst: Colleen O'Bryant

1. Reason for instrument evaluation

Instrument unresponsive, screen black.

2. Visual Inspection

3. Troubleshooting/Repair

Replaced the power supply, PI board and controller board <sup>and</sup> individually and the instrument is still unresponsive.

Called Intel, spoke to Cindy, she requested the instrument be sent to them.

4. VOC (attached) NA

5. Internal Standard value  $\leq 3.00\%$  Yes  No  NA

6. Verify software version  NA

Return to Factory for Repair

Ready For Calibration

Ready For Certification

# Repair

**Intoximeters**  
Experience • Service • Integrity

You are receiving this because you recently contacted Intoximeters and your instrument(s) needs to be returned for repair. Please review the top portion of this form for accuracy, and complete the bottom portion. Include a completed copy of this form with your shipment for faster service.

Incident #: **INC-019123**

Description: **DMT / Dead**

Customer: Alaska Dept Of Public Safety

Contact: Colleen OBryant

Phone: 907-561-1092

Email: Colleen.obryant@alaska.gov

**Line**      **Unit**

1      LEGDMT

**Serial Number**

50100689

General Reason: DMT Func

Specific Reason:

Note:

Customer changed the Power supply, the Control board and the DSI Board. The board is lit up green, so it does have power. She's run out of parts, and ideas.

Please ship your item to :

**Intoximeters Inc.**  
**2081 Craig Road St. Louis, MO 63146 USA**

Please complete this section :

Shipping : Please select the method for return shipping to you : (choose one)

- Next Day     2nd Day     Ground     Collect (if collect, please provide account number)

Authorization : (chose one)

- I authorize all repairs     I authorize all repairs up to \$ 100  
 I require a written estimate prior to repair completion (a \$40.00 estimate fee applies)

Calibration :

- I require an ISO 17025 accredited certificate of calibration. \*Not available on all models.  
Fees apply: \$50/Hand-held and True Cal, \$75/Desktop

Payment Method : (chose all that apply)

- No purchase order required -- work authorized by: \_\_\_\_\_  
 A purchase order number is required before repair of instrument. PO# 16-1024003  
 A PHYSICAL PO is required. (**Hospitals must provide**) Please submit along with this form.  
 Pay by Credit Card -- Please provide the following:

Credit Card Number \_\_\_\_\_ Name on the Card \_\_\_\_\_  
Expiration Date: \_\_\_\_\_ CVV2 Number: \_\_\_\_\_  
Billing Address \_\_\_\_\_

Address Confirmation : (correct if necessary)

Billing Address :

Alaska Dept Of Public Safety  
524 E 48th Ave  
Anchorage AK 99503  
USA

Shipping Address :

Alaska Dept Of Public Safety  
524 E 48th Ave.  
Anchorage AK 99503  
USA

*Alaska Scientific Crime Detection Laboratory  
4805 Dr. MLK Jr. Ave  
Anchorage, AK 99503 99504*