

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100699

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 03/12/2018

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.54: 0.079
LOT #: AG735001
EXPIRATION: 12/16/2019
TANK PRESSURE: 1180 psi

| | | |
|-------------------|----------|-------|
| BLANK TEST | 0.000 | 20:44 |
| INTERNAL STANDARD | VERIFIED | 20:45 |
| EXTERNAL STANDARD | 0.077 | 20:45 |
| BLANK TEST | 0.000 | 20:46 |
| EXTERNAL STANDARD | 0.077 | 20:46 |
| BLANK TEST | 0.000 | 20:47 |
| EXTERNAL STANDARD | 0.077 | 20:47 |
| BLANK TEST | 0.000 | 20:48 |
| EXTERNAL STANDARD | 0.077 | 20:48 |
| BLANK TEST | 0.000 | 20:49 |
| EXTERNAL STANDARD | 0.077 | 20:49 |
| BLANK TEST | 0.000 | 20:50 |

Average = 0.0770
Std Dev = 0.0000

Diagnostic Check

VERSIONS
DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 49.1°C PASSED
Breath Tube = 42.6°C PASSED

PUMP INFO

Flow Rate = 4.256 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED


FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD PASSED


I, Charles R. Foster, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Charles R. Foster
Scientific Director
State Breath Alcohol Program

5/3/18

Subscribed and sworn before me this 03 day of 05, 2018


Carlie K. Bailey, Notary Public
My Commission Expires With Office



0034/25/18