

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100681

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 01/03/2026

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080

TARGET AT 29.52: 0.079

LOT #: AG436502

EXPIRATION: 12/30/2026

TANK PRESSURE: 631 psi

BLANK TEST	0.000	17:20
INTERNAL STANDARD	VERIFIED	17:20
EXTERNAL STANDARD	0.078	17:20
BLANK TEST	0.000	17:21
EXTERNAL STANDARD	0.077	17:21
BLANK TEST	0.000	17:22
EXTERNAL STANDARD	0.078	17:22
BLANK TEST	0.000	17:23
EXTERNAL STANDARD	0.078	17:23
BLANK TEST	0.000	17:24
EXTERNAL STANDARD	0.078	17:25
BLANK TEST	0.000	17:25

Average = 0.0778

Std Dev = 0.0004

Diagnostic Check

VERSIONS

DMT: 3.02

PIC: 3.03

Modem: 2.6

Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C PASSED

Breath Tube = 45.6°C PASSED

PUMP INFO

Flow Rate = 4.282 L/M PASSED

DETECTOR INFO

PUMP ON PASSED

PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED

Filter 2 PASSED

Filter 3 PASSED

INTERNAL STANDARD PASSED

File Only
CHF 2/2/26