

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Issued: 6/11/2014  
Effective: 6/11/2014

Version: CIISF 2014 R0  
Status: Active

Date: 6/24/15 Instrument Serial Number: 100384

Instrument Location: BARROW - POLICE - SQUAD

Supervisor Name and Agency: SGT JAMIES MITCHELS 

Describe the Status Change or Issue with the Instrument:

CONTINUED INTERVAL STANDARD ERRORS - 4 IU  
SEVERAL DAYS. NEW INSTRUMENT BEING SET UP

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: \_\_\_\_\_

Out of Service Date/Time: 6/24/15 9:58 AM

For Use by SCDL

Additional Notes

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)

For questions contact the Breath Alcohol Program at 907-269-5740