

# Alaska Scientific Crime Detection Laboratory

## DataMaster Evaluation Form

Effective: 10/10/2022

Version: 3.0

Instrument #: 100684  
Start Date: 12/14/22  
Date Received: 12/14/22

Analyst(s): C. OByant

Reason for instrument evaluation

undefined error  
unable to load ticket db

Visual Inspection

- |  |  |
|--|--|
| <input type="checkbox"/> Clean   | <input type="checkbox"/> Replace missing screws            |
| <input type="checkbox"/> Cermetex Box (modem) is secure                            | <input type="checkbox"/> Ensure modem cable present        |
| <input type="checkbox"/> Double Check Valve secured to wall                        | <input type="checkbox"/> All tubing and connections secure |
| <input type="checkbox"/> O-ring inspection and ensure 2 replacement O-rings in bag |  |

As Found Conditions:

Barometric Pressure (inHg): \_\_\_\_\_ Temperature (C): \_\_\_\_\_ Relative Humidity (%) \_\_\_\_\_

Linearity Test (attached)

0.020g/210L Mnf: \_\_\_\_\_ Lot: \_\_\_\_\_ Exp: \_\_\_\_\_

0.080g/210L Mnf: \_\_\_\_\_ Lot: \_\_\_\_\_ Exp: \_\_\_\_\_

0.300g/210L Mnf: \_\_\_\_\_ Lot: \_\_\_\_\_ Exp: \_\_\_\_\_

Notes

Unable to perform as-found due to ticket db  
pop-up message.

- |   |   |
|---|---|
| <input type="checkbox"/> VOC (attached)                           | <input type="checkbox"/> Verify software version              |
| <input type="checkbox"/> VOC Internal Standard value $\leq$ 3.00% | <input type="checkbox"/> Non-Drinking Subject Test (attached) |

Condition of item:  Good  Needs Repair

- Ready for Adjustment  
 Ready for Calibration

End Date: 12/14/22