

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 4.0

Effective: 8/10/2022

Date: 7/13/2023

Instrument #: 100683

Instrument Location: Ketchikan, AK

Supervisor Name and Agency: Trooper Charles Strickland

Describe the Status Change or Issue with the Instrument:

Old DMT #100398 had a filter wheel error and is being taken out of service. New DMT #100683 has been put into service.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: 7/13/23, 1200 hours

Out of Service Date/Time: *Click or tap here to enter text.*

For Use by SCDL

Additional Notes

Click or tap here to enter text.

Email completed form to dps.scdl.toxicology@alaska.gov
For questions contact the Breath Alcohol Program at 907-269-5740

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor