

# Alaska Scientific Crime Detection Laboratory

## DataMaster Check In Form

Issued: 2/2/2015  
Effective: 2/2/2015

Version: DMCIF 2015 R0  
Status: Active

Instrument S/N 100411

Received From Keno: PD

Date 2/27/15

Analyst CTCF

1. Reason for instrument return:  
Interference detected on <sup>CW</sup> external standard and internal standard errors
2. Change instrument and case location in JusticeTrax Breath Alcohol Database
3. Ensure change in instrument status form is in JusticeTrax
4. Visual Inspection
5. Upload
6. Update agency information and polling status in DM Host Calling Program
7. Check for VOCs

From NPAS

8. NPAS Repair Paperwork (attached)