

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Issued: 7/10/2015
Effective: 7/10/2015

Version: DMCIF 2015 R1
Status: Active

Instrument S/N 100677

Received From Bethel PD

Date 6/13/16

Analyst Colleen O'Bryant

1. Reason for instrument return:

Removed from service for various messages (undefined error and unable to load ticket). Upon return DMT showing similar messages @ SCOT. Unable to upload or put files on USB

2. Change instrument and case location in JusticeTrax Breath Alcohol Database

3. Ensure change in instrument status form is in JusticeTrax

4. Upload (unable to upload)

5. Update agency information and polling status in DM Host Calling Program

6. Check for VOCs (unable to upload)

From NPAS

7. NPAS Repair Paperwork (attached)