

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument  
State of Alaska

Serial #: 100669

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 07/05/2015

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 30.18: 0.081  
LOT #: 01315080A3  
EXPIRATION: 02/05/2017  
TANK PRESSURE: 1108 psi

|                   |          |       |
|-------------------|----------|-------|
| BLANK TEST        | 0.000    | 00:02 |
| INTERNAL STANDARD | VERIFIED | 00:02 |
| EXTERNAL STANDARD | 0.081    | 00:03 |
| BLANK TEST        | 0.000    | 00:04 |
| EXTERNAL STANDARD | 0.081    | 00:04 |
| BLANK TEST        | 0.000    | 00:05 |
| EXTERNAL STANDARD | 0.081    | 00:05 |
| BLANK TEST        | 0.000    | 00:06 |
| EXTERNAL STANDARD | 0.081    | 00:06 |
| BLANK TEST        | 0.000    | 00:07 |
| EXTERNAL STANDARD | 0.081    | 00:07 |
| BLANK TEST        | 0.000    | 00:08 |

Average = 0.0810  
Std Dev = 0.0000

## Diagnostic Check

### VERSIONS

DMT: 3.01  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.9°C PASSED  
Breath Tube = 46.7°C PASSED

### PUMP INFO

Flow Rate = 4.637 L/M PASSED

### DETECTOR INFO

PUMP ON PASSED  
PUMP OFF PASSED

### FILTER INFO


Filter 1 PASSED  
Filter 2 PASSED  
Filter 3 PASSED

INTERNAL STANDARD

PASSED

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

8/7/2015

Subscribed and sworn before me this 7<sup>th</sup> day of Aug., 20 15

  
Nikki Roth, Notary Public  
My Commission Expires With Office



COB 7/30/15