

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100689

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 12/27/2022

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 28.99: 0.078  
LOT #: AG113402  
EXPIRATION: 05/14/2023  
TANK PRESSURE: 371 psi

BLANK TEST	0.000	08:54
INTERNAL STANDARD	VERIFIED	08:54
EXTERNAL STANDARD	INTERFERENCE	08:54

Blank Error

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.07  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.9°C	PASSED
Breath Tube = 38.4°C	PASSED

### PUMP INFO

Flow Rate = 4.083 L/M	PASSED
-----------------------	--------

### DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

### FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------

File Only jhw 02/10/23