

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100414

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 10/24/2020

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 28.99: 0.078  
LOT #: AG923401  
EXPIRATION: 08/22/2021  
TANK PRESSURE: 995 psi

BLANK TEST                      0.000    12:02  
INTERNAL STANDARD            VERIFIED 12:02

Gas Flow Error

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.02  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 49.0°C            PASSED  
Breath Tube       = 48.1°C           PASSED

### PUMP INFO

Flow Rate = 4.683 L/M               PASSED

### DETECTOR INFO

PUMP ON                                PASSED  
PUMP OFF                               PASSED

### FILTER INFO

Filter 1                                 PASSED  
Filter 2                                 PASSED  
Filter 3                                 PASSED

### INTERNAL STANDARD

PASSED