

# Alaska Scientific Crime Detection Laboratory

## DataMaster Check In Form

Effective: 10/10/2022

Version: 3.0

Instrument #: 100357

Received From: Wrangell PD

Date: 8/1/2024, 9/4/2024

Analyst: Megan Carlisle

COB 2024.10.07

Colleen O'Bryen

Reason for instrument return:

Instrument experienced intermittent Interference Detected messages.  
Instrument experienced unable to load ticket database status message  
and became unresponsive.

Change instrument and case location in JusticeTrax Breath Alcohol Database

Ensure change in instrument status form is in JusticeTrax

Upload  \* COB 9/4/24

Update agency information and polling status in DMHost Calling Program

Check for VOCs  COB 2024.09.04

### From Intoximeters

Intoximeters Repair Paperwork (attached)

\* Ticket db was corrupt + causing software to shut down + restart instrument. Navigated to windows explorer in DMT software to delete corrupt ticket db file via the SD card. Replaced SD card in the DMT and restarted numerous times. uploaded successfully. COB 9/4/24