

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: April 26, 2017 Instrument Serial Number: 100397

Instrument Location: Fairbanks

Supervisor Name and Agency: Elondre Johnson Alaska State Troopers

Describe the Status Change or Issue with the Instrument: The instrument was taken out of service
because the crime lab was not able to read the reports from the instrument

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: April 26, 2017 / 1430

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740

O'Bryant, Colleen S (DPS)

From: O'Bryant, Colleen S (DPS)
Sent: Tuesday, March 07, 2017 2:25 PM
To: Johnson, Elondre A (DPS)
Cc: Roth-Frey, Nicolette A (DPS); Foster, Charles R (DPS)
Subject: AST/Fairbanks DMT

Elondre,

After unsuccessfully troubleshooting your DMT with the local fax line, I recommend we replace your instrument. Please fill out the appropriate paperwork to remove the instrument from service and we will mail you a new one.

Thanks for your patience and help.

Colleen