

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 6/10/16 Instrument Serial Number: 100370

Instrument Location: Homer PD

Supervisor Name and Agency: Will Huff (Per formed) Brandi Barnett-SCDL

Describe the Status Change or Issue with the Instrument:

We were notified by the current supervisor at Homer PD that after changing the external standard the tank was leaking. He tried taking it in and out but the pressure was still dropping. We provided directions for changing the o-ring on the regulator as well as the specifications for the o-ring type and Will Huff (a breath test operator) replaced the o-ring on 6/10/16 with no further drops in tank pressure observed. The maintenance was performed by Will Huff at the direction of the crime lab as the supervisor was unavailable.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: _____

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740