

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100689

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 10/30/2018

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.91: 0.080
LOT #: AG735001
EXPIRATION: 12/16/2019
TANK PRESSURE: 920 psi

| | | |
|-------------------|----------|-------|
| BLANK TEST | 0.000 | 19:43 |
| INTERNAL STANDARD | VERIFIED | 19:43 |
| EXTERNAL STANDARD | 0.075 | 19:43 |

Blank Error

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

| | |
|-------------------------|--------|
| Sample Chamber = 48.9°C | PASSED |
| Breath Tube = 41.4°C | PASSED |

PUMP INFO

| | |
|-----------------------|--------|
| Flow Rate = 4.482 L/M | PASSED |
|-----------------------|--------|

DETECTOR INFO

| | |
|----------|--------|
| PUMP ON | PASSED |
| PUMP OFF | PASSED |

FILTER INFO

| | |
|----------|--------|
| Filter 1 | PASSED |
| Filter 2 | PASSED |
| Filter 3 | PASSED |

INTERNAL STANDARD

PASSED