

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 10/15/19 Instrument Serial Number: 100391

Instrument Location: Soldotna AST

Supervisor Name and Agency: Ronny Simmons / AST

Describe the Status Change or Issue with the Instrument:

Took this DataMaster out of service
and packaged to be sent to lab.

Vac completed prior to taking out of service

~~Time was off 1 hr of set up so it was~~
~~pretty off 2 hrs AST~~

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 10/15/19 / ~~10/15~~ AST 0911

For Use by SCDL

Additional Notes

Internal standard errors, tan interference from
In field Instrument Review. JMW 10/21/19

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740