

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Effective: 10/10/2022

Version: 3.0

Instrument #: 100356

Received From: Whittier-Girdwood PD

Date: 9/27/24 - 10/2/24

Analyst: Megan Carlock

Reason for instrument return:

Received a fatal error and was dropped off without a keyboard. When instrument was turned on an error message "unable to load ticket database" occurred.

Change instrument and case location in JusticeTrax Breath Alcohol Database

Ensure change in instrument status form is in JusticeTrax

Upload

Update agency information and polling status in DMHost Calling Program

Check for VOCs

From Intoximeters

Intoximeters Repair Paperwork (attached)

Derek Watten retrieved data files from SD and loaded into DMHost on 10/02/24. One VOC was located (09/05/24). SD card was replaced. DM 10/02/24